

Technical Evaluation report against tender ref. no.-UPMSCL/EQ/RC/320 for Radiant Warmer

Sr. no.	Name of the Schedule	Conclusion of the received bids
1.	Radiant Warmer	<p>Qualified Bids</p> <ul style="list-style-type: none">a. M/s. Phoenix Medical System Pvt. Ltd.b. M/s. Medilux Systemc. M/s AVI Healthcare <p>Disqualified Bids</p> <ul style="list-style-type: none">d. M/s Medisense Electronics <p>Reasons:</p> <ul style="list-style-type: none">1. Not Submitted Quality certificates like: USFDA or European CE, IEC-60601-1-2, IEC 60601-2-212. Not submitted Performance certificate for same quoted model (atleast 10% of the schedule quoted).

N.B.: If there is any representation against above mentioned technical evaluation report kindly submits it by 27.03.2021 by 12:00 P.M., any representation submitted after above mentioned time period will not be entertained.