



UTTAR PRADESH MEDICAL SUPPLIES CORPORATION LTD.

Regd. Office: SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010

Ph: 0522 2838102 e-mail: drugs@upmsc.in

Result of Technical and Sample Evaluation of Tender no. UPMSCL/Drugs-112/459

Dated: 14.07.2021

S. No	Name of Firm	Technical ly Qualified / Not qualified	Reason for technical disqualification as on 25.06.2021	Sample submission status	Sample evaluation result	Reason for Sample disqualification	Final result (based on Technical and Sample Evaluation)
1	M/s Hi Care Gloves Pvt Ltd	Not-Eligible	<p>-MSME and DIC certificate are not legible.</p> <p>-Firm has not submitted copy of Quality compliant with the below standards, or equivalent:</p> <p>- EU standard directive 93/42/EEC Class I, EN 455</p> <p>- EU standard directive 89/686/EEC Category III, EN 374</p> <p>- ANSI/SEA 105-2011</p> <p>- ASTM D6319-10</p> <p>-Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS13422) for Hi care Gloves Pvt Ltd & submitted IS13422 is issued for Neerakkal Gloves Pvt Ltd</p> <p>-Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS 15354).</p> <p>-Firm has not submitted 60 days production capacity for nitrile gloves</p> <p>-Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves and rubber examination gloves.</p> <p>-Firm has not submitted average annual turnover certificate with audited /provisional balance sheet for the FY 2020-21.</p>	Submitted	<p>Recommended for Rubber examination gloves</p> <p>Recommended for Sterilized Surgical Gloves Pair</p> <p>Not Recommended for Nitrile Gloves</p>	<p>-</p> <p>-</p> <p>All size samples are not submitted</p>	Not eligible
2	M/s Jajoo Surgicals Pvt. Ltd.	Not-Eligible	<p>-Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS 15354).</p>	Submitted	<p>Not recommended for Nitrile Gloves</p> <p>Recommended for Rubber examination gloves</p>	<p>Short length & wearing purpose</p> <p>-</p>	Not eligible
3	M/s M/s Sanmati Packaging	Not-Eligible	<p>-Firm has not submitted copy of Quality compliant with the below standards, or equivalent:</p> <p>- EU standard directive 93/42/EEC Class I, EN 455</p> <p>- EU standard directive 89/686/EEC Category III, EN 374</p> <p>- ANSI/SEA 105-2011</p> <p>- ASTM D6319-10</p> <p>-Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS 15354).</p> <p>-Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS13422).</p>	Submitted	<p>Nitrile Gloves</p> <p>Recommended for Rubber examination gloves</p>	<p>Not Recommended (All size samples are not submitted)</p> <p>-</p>	Not eligible

S. No	Name of Firm	Technically Qualified / Not qualified	Reason for technical disqualification as on 25.06.2021	Sample submission status	Sample evaluation result	Reason for Sample disqualification	Final result (based on Technical and Sample Evaluation)
					Not Recommended for Sterilized Surgical Gloves Pair	Sample is submitted of other manufacturer as observed on packing	
4	M/s National Meditek	Not-Eligible	-Firm has not submitted average annual turnover certificate with audited /Provisional balance sheet for the FY 2020-21.	Submitted	Not Recommended for Nitrile Gloves	All size samples are not submitted	Not eligible
5	M/s RFB Latex Limited	Not-Eligible	Firm has not submitted copy of Quality compliant with the below standards, or equivalent: - EU standard directive 93/42/EEC Class I, EN 455 - EU standard directive 89/686/EEC Category III, EN 374 - ANSI/SEA 105-2011 - ASTM D6319-10 -Firm has not submitted copy of test report from Government laboratory/NABL for sterile surgical gloves pair.	Submitted	Not recommended for Nitrile Gloves Recommended Sterilized Surgical Gloves Pair	Short length -	Not eligible
6	M/s Sky Lifecare Pvt Ltd	Not-Eligible	-As per tender specification, Outer gloves preferably reach mid-forearm (minimum 280 mm total length) while firm is quoted for 240 mm only. -Firm has not submitted copy of Quality compliant with the below standards, or equivalent: - EU standard directive 93/42/EEC Class I, EN 455 - EU standard directive 89/686/EEC Category III, EN 374 - ANSI/SEA 105-2011 - ASTM D6319-10 -Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS 15354). -Firm has not submitted copy of authorization letter from Importer. -Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves and rubber examination gloves. -Firm has not submitted average annual turnover certificate with audited /provisional balance sheet for the FY 2020-21.	Submitted	Not recommended for Nitrile Gloves Recommended for Rubber examination gloves	Easily breakable on slight pulling. -	Not Eligible
7	M/s Sogo Fashions Private Limited	Not-Eligible	-	Submitted	Not recommended for Nitrile Gloves	Easily breakable on slight pulling	Not Eligible
8	M/s Unity Healthcare	Not-Eligible	-	Submitted	Not recommended for Nitrile Gloves	Easily breakable on slight pulling	Not Eligible

Managing Director

UPMSCL