

Technical Evaluation Report against Tender Ref. no.-UPMSCL/EQ/RC/155 floated for Procurement of Electrolyte Analyzer.

Sr. no.	Name of the Schedule	Conclusion of the received bids
1.	Electrolyte Analyzer	<p>Qualified bids: None</p> <p>Disqualified: 1. M/s. Lab X. Reasons: a. The quoted model does not support ca+ test.</p> <p>2. M/s. Sensacore Reasons: a. Disqualified as they have not submitted USFDA.</p> <p>3. M/s. Transasia Reasons: a. Disqualified as the throughput for urine is 90 sec and the required throughput should be 60 sec.</p> <p>4. M/s. POCT. Reasons: a. The quoted model does not support ca+ test.</p> <p>5. M/S Surbhi Meditek Reasons: a. The quoted model does not support ca+ test.</p>

N.B.:

A. If there is any representation against uploaded Technical Evaluation Report kindly submits the same by 31.12.2020 till 3:00 P.M.

B. Tentative date for Physical Demonstration of all qualified bidders against above mentioned item is 06.01.2021 and final date will be intimate through email.