



UTTAR PRADESH MEDICAL SUPPLIES CORPORATION LTD.
(A Government of Uttar Pradesh Undertaking)

Regd. Office: SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010

Corrigendum-3 dated 15.02.2021

With reference to tender no. **UPMSCL/Drugs-090/327** dated 19 January, 2021 a corrigendum is being issued as follows:

S.N.	Annexure-A Schedule of requirement	Existing	Revised/ clarification
1	RDT (HEPA-C) Rapid Diagnostic Kit For Hepatitis-c	Kit needs to be able to detect both antibodies (capsid, NS3 and NS4) and antigen (Capsid) against/of Hepatitis C Virus.	Kit needs to be able to detect antibodies (capsid, NS3 and NS4) against/ of Hepatitis C virus.
2		The assay should have sensitivity of > 99 % and specificity of > 99%.	The assay should have sensitivity of more than or equal to 99% and specificity of more than or equal to 98%.
3	ELISA KIT (HEPA-C) ELISA Kit for Hepatitis-C	Kit needs to be able to detect both antibodies (capsid, NS3 and NS4) and antigen (Capsid) against/of Hepatitis C virus.	Kit needs to be able to detect antibodies (capsid, NS3 and NS4) against/ of Hepatitis C virus
4		The assay should have sensitivity of > 99 % and specificity of > 99%.	The assay should have sensitivity of more than or equal to 99% and specificity of more than or equal to 98%.
5	ELISA KIT- (HIV) ELISA kit for HIV	The assay should detect HIV 1 & 2 antibodies.	The assay should detect both HIV 1 & 2 antibodies and antigen.

All other terms and conditions of the tender document shall remain same.

MANAGING DIRECTOR